amc technical brief

Editor M Thompson

Analytical Methods Committee AMCTB No 16

Revised April 2007

Proficiency testing: assessing z-scores in the longer term

While a single z-score provides a valuable indication of the performance of a laboratory, a set or sequence of z-scores provides a deeper insight. Both graphical and numerical methods may be appropriate for assessing a sequence of zscores. However, due caution is required with numerical methods to avoid incorrect conclusions. The use of a summary score derived from z-scores relating to a number of different analytes is not recommended: it has a very limited range of valid applications and tends to conceal sporadic or persistent problems with individual analytes. Moreover it is prone to misuse by non-scientists.

Summary scores

The following two types of summary score are statistically soundly based [1] and may be useful for individual participants to assess a sequence of z-scores z_i , i = 1,...,n derived from a single combination of analyte, test material and method.

The rescaled sum of the z-scores,

$$RSZ = \sum_{i} z_{i} / \sqrt{n} ,$$

can be interpreted on the same basis as a single z-score, i.e., it is expected to be zero-centred with unit variance if the z-scores are. This statistic has the useful property of demonstrating a persistent bias or trend, so that the sequence of results [1.5, 1.5, 1.5, 1.5] provides a statistically significant RSZ of 3.0, even though any single one result is not significant at the 95% confidence level. However, RSZ could conceal two large zscores of opposite sign that roughly cancel.

The sum of the squared z-scores,

$$SSZ = \sum_{i} z_i^2 ,$$

could be interpreted as a χ_n^2 distribution for zero-centred zscores with unit variance. This statistic has the advantage of avoiding the cancellation of large z-scores of opposite sign, but is less sensitive to small biases. SSZ is especially sensitive to

Both of these summary statistics need to be protected (for example, by robustification or filtering) against past outlying scores, which would otherwise have a long-term persistence

Control charts

Graphical methods of summarising a set of z-scores can be just as informative as summary scores and are probably less prone to misinterpretation. Multiple univariate charts [2], such as those shown overleaf, give a clear overview and are especially useful when scores from a group of analytes determined by a common method are considered. Hand-drawn charts are quick to update and serve just as well as those produced by computer.

The control chart (Figure 1 overleaf) shows upward-pointing symbols to indicate z-scores greater than zero and downwardpointing symbols for those less than zero. Small symbols represent instances where $2 \le |z| < 3$, and large symbols

instances where $|z| \ge 3$. The data illustrated immediately show some noteworthy features. Results from round 11 are mostly too low, demonstrating a procedure that was faulty in some general feature, while analyte 7 gives high results too frequently, demonstrating a persistent problem with that specific analyte. The remaining results are roughly consistent with fitness for purpose, which on average would result in about 5% of z-scores

J-charts

A J-chart (otherwise known as a 'zone chart') [3] is even more informative, because it combines the capabilities of the Shewhart and the cusum charts. It does this by cumulating special J-scores attributed to successive results on either side of the zero line. This enables persistent minor biases to be detected as well as abrupt large changes in the analytical system. Typical rules for converting z-scores to J and cumulating them are as follows.

• If $2 \le |z| < 3, \quad J = 4.$ $1 \le |z| < 2, \quad J = 2.$

represented by a small symbol.

- J-scores from successive rounds are cumulated until $J \ge 8$, which defines an excursion beyond the action limits and triggers investigative procedures.
- The cumulator is reset to zero after any such excursion, before resuming cumulation.
- The cumulator is reset to zero, before resuming cumulation, when the new value of z is of opposite sign to the previous value or reported as exactly zero.

Examples

Successive values of z go from left to right.

Z	1.5	1.5	-0.9	-1.5
J	2	2	0	2
Cumulator	2	4	0	2
Z.	1.5	1.5	-1.5	-1.5

z	1.5	1.5	-1.5	-1.5
J	2	2	2	2
Cumulator	2	4	2	4

It is an optional aid to add a minus sign to J-scores resulting from negative z-scores thus:

z	1.5	1.5	-1.5	-1.5
J	2	2	-2	-2
Cumulator	2	4	-2	-4

Several examples of the cumulative effect of bias are visible in Figure 2 (which illustrates the same results as Figure 1 for comparison). For example, Analyte 3 in Rounds 1-4 receives z-scores of 1.5, 1.2, 1.5, and 1.1 respectively, translating into J-values of 2, 2, 2, and 2, which cumulate to 8 by Round 4 and trigger investigative procedures.

References

- 1. M Thompson, S L R Ellison and R Wood, *Pure Appl Chem*, 2006, **78**, 145-196.
- 2. M Thompson, K M Malik and R J Howarth, *Anal Comm*, 1998, **35**, 205-208.
- 3. AMC Technical Briefs, No 12 (2003).




Figure 1. A multiple control chart for z-scores.




Figure 2. A multiple J-chart for z-scores (same data as Figure 1). Note that negatively-signed J-scores are used in this chart. (Examples of standard (univariate) J-charts are shown in Reference 3.)

Other AMC Technical Briefs can be found on: www.rsc.org/amc